

Summary Calibration Report

Ace Pipettes, LLC.
 14230 Prospect Point Drive
 Cypress, TX, 77429
 832-295-0882

Date: 07/29/2019
 Work Order: KB0729191
 Company/Inst: MD Anderson Cancer Center
 Dept: Pediatrics Research
 Name: Simon Olivares
 PO:

Serial#	ID1	ID2	Volume "UL"	Make	Model	Cal. Date	Due Date	As Found	As Left	Type	
1	17607987	FLUIDIGM STATION	1000	Sartorius	Biohit Mline	07/29/2019	07/28/2020	NA	Pass	Single	
2	19004804	FLUIDIGM STATION	100	Sartorius	Biohit Mline	07/29/2019	07/28/2020	NA	Pass	Single	
3	17596299	FLUIDIGM STATION	10	Sartorius	Biohit Mline	07/29/2019	07/28/2020	NA	Pass	Single	
4	17571680	FLUIDIGM STATION	3	Sartorius	Biohit Mline	07/29/2019	07/28/2020	NA	Pass	Single	
5	16528797	FLUIDIGM STATION	10	Sartorius	Biohit ProLine Plus	07/29/2019	07/28/2020	NA	Pass	8Ch	
6	18087749	FLUIDIGM STATION	10	Sartorius	Biohit Mline	07/29/2019	07/28/2020	NA	Pass	12Ch	
7	KH21724	SCR 1.2158 HOOD 1	1000	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
8	KH26917	SCR 1.2158 Counter	1000	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
9	DH38224	SCR 1.2158 HOOD 3	1000	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
10	AA94321	SCR 1.2158 HOOD 2	1000	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
11	KH35074	SCR 1.2158 Counter	100	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
12	KH35065	SCR 1.2158 HOOD 1	100	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
13	CH25692	SCR 1.2158 HOOD 3	100	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
14	EJ02258	SCR 1.2158 HOOD 2	100	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
15	CH30934	SCR 1.2158 HOOD 3	10	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
16	DH34993	SCR 1.2158 HOOD 1	10	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
17	CJ16651	SCR 1.2158 HOOD 3	10	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
18	840250029		200	lsc BioExpress	GeneMate	07/29/2019	07/28/2020	NA	Pass	Single	
19	AA94344	SCR 1.2116	PCR HOOD	1000	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single
20	DH17672	SCR 1.2158 HOOD 1	1000	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
21	CH12393	SCR 1.2110	STATION#16	300	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single
22	CJ11159	SCR 1.2114	PCR HOOD RNA	100	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single
23	DH75442		50	Thermo	Finnpipette	07/29/2019	07/28/2020	NA	Pass	Single	
24	CH16991	SCR 1.2116	PCR HOOD gDNA	10	Thermo	Finnpipette F1	07/29/2019	07/28/2020	NA	Pass	Single
25	MU34067		1000	Thermo	Finnpipette F1	07/29/2019	07/28/2020	NA	Pass	Single	
26	GH18961	SCR 1.2110	STATION#16	100	Thermo	Finnpipette F1	07/29/2019	07/28/2020	NA	Pass	Single
27	MU34144		100	Thermo	Finnpipette F1	07/29/2019	07/28/2020	NA	Pass	Single	
28	GH09875	SCR 1.2110	STATION#16	10	Thermo	Finnpipette F1	07/29/2019	07/28/2020	NA	Pass	Single
29	MU34116		10	Thermo	Finnpipette F1	07/29/2019	07/28/2020	NA	Pass	Single	
30	950070028	SCR 1.2158	200	lsc BioExpress	GeneMate	07/29/2019	07/28/2020	NA	Pass	12Ch	
31	751170023	SCR 1.2158	200	lsc BioExpress	GeneMate	07/29/2019	07/28/2020	NA	Pass	12Ch	
32											
33											
34											
35											
36											
37											
38											
39											
40											
41											
42											
43											
44											
45											

Number of pipets:	Single Channel	Multi Ch "8"	Multi Ch "12"
Level 2	27	1	3

Calibrated by: KB
 Balance/s: Recal due:
 B043079758 OCT/31/2018

Procedure/Notes:
 These pipettes were calibrated gravimetrically with water that is no less than 17 MegaOhm/cm3.
 All balances are calibrated against NIST traceable standards and serviced yearly by the manufacturer.
 This document cannot be reproduced partially.